



Headoffice: C-4th&5th, 16, Deogyong-daero 1556beon-gil, Yeongtong-gu, Suwon-si, Gyeonggi-do, 16690, REPUBLIC OF KOREA
Manufacturing site: 74, Osongsaengmyeong 4-ro, Osong-eup, Heungdeok-gu, Cheongju-si, Chungcheongbuk-do, 28161, REPUBLIC OF KOREA

TEL : +82-31-300-0400 / FAX : +82-31-300-0499
<http://www.sdbiosensor.com>

Certificate of Analysis

Product Name	SARS-CoV-2 Rapid Antigen Test		
Cat No.	99COV30D	Ref No.	9901-NCOV-01G
Lot No.	QCO391039I	MFG Date	2021.01.13.
		EXP Date	2023.01.12.
Testing categories	Specification	Result	
Apperance Inspection	<ul style="list-style-type: none"> • The membrane is well seated • The mark on the appearance of the device is correct. • Extraction buffer is a colorless liquid formulation • All Desiccant color must be yellow 	Pass	
Sensitivity	In-house reference panels for the sensitivity test 1) All panels must be interpreted as positive.	Pass	
Specificity	In-house reference panels for the specificity test 1) The test lot specificity should not be less than 96.5% (95% CI, 91.33%-99.04%).	Pass	
Migration test	The migration of the product should appear in the results window within 2 minutes and should reach the end of the results window within 6 minutes.	Pass	
Background test	The background of the product should be light red or clean within the Reading time.	Pass	
Packaging Inspection	Strip Devices (25 strips) / Sterile swab (25ea) / Extraction buffer (25ea) / Insert (1ea) / Nozzle cap (25ea) / Clear film (1ea) / Quick Reference Guide (1ea)	Pass	
	Manufacturing and expiration date between the pouch and package of test kit	Pass	
	LOT No. between the pouch and package of test kit	Pass	

According to SD Biosensor's quality assurance system, the above QC tests have been completed and passed the specification. We hereby approve this product release.

SD Biosensor, Inc

Geunkuk Song
 Quality Management Representative
 QA Division

Geunkuk Song
 Quality Management Representative
 QA Division